

April 4, 1966

Grass Instrument Company  
101 Old Colony Avenue  
Quincy, Massachusetts

Gentlemen:

Separately we are shipping you one oscillograph, Serial No. G331V5, Model 6A, from our electroencephalograph. The write out with the enclosed oscillograph has a response time delay which looks like a prolonged time constant. Replacing the enclosed oscillograph with a known good unit eliminated the problem.

Would you please either replace or repair this unit? This is to be charged to the University of Missouri, and we will issue the appropriate repair orders when you send us the charges.

Sincerely yours,

Max Fink, M.D.  
Professor of Psychiatry

MF:kp